

### Certificate of Analysis

**Product: Hydrogen Peroxide 3% Solution**

**Item Number: HPL3**

**Grade: Lab**

**Lot Number: 5001473/1.1-3-102925**

**Manufacture Date: 10/29/2025**

**Expiration Date: 10/28/2027**

**Country of Origin: United States**

Tested Property	Specification	Analysis
H2O2 Assay	3.00 - 3.50 vol%	3.00
Water	96.50 - 97.00 vol%	97.00
Appearance	Clean, Water White	Pass
Specific Gravity @ 25°C	As reported	1.0105

#### [Hydrogen Peroxide 3% Solution](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,  
E-mail: [customercare@laballey.com](mailto:customercare@laballey.com),  
[www.laballey.com](http://www.laballey.com),

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2026 Lab Alley