

Certificate of Analysis

Product: Water Purified
Item Number: WAU
Grade: USP
Lot Number: U-10626-D
Manufacture Date: 01/06/2026
Expiration Date: 01/05/2028
Country of Origin: United States

Tested Property	Specification	Analysis
Appearance	Clear, colorless liquid	Passes
Water Conductivity (Point of fill)	$\leq 1.3 \mu\text{S/cm}$ at 25°C	0.22
Microbial Limits	≤ 100 CFU/mL	<1
Total Organic Carbon	<0.50 mg C/L	0.01

Due to change in USP monographs for purified water, pH and permanganate test are no longer required to be tested and are not included on the C of A.

[Water Purified](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2026 Lab Alley