

Certificate of Analysis

Product: Chloroform \geq 99% Lab Grade

Item Number: CHLL

CAS#: 67-66-3

Grade: Lab

Lot Number: 5002180/1.1

Manufacture Date: 04/18/2026

Expiration Date: 04/17/2029

Country of Origin: United States

Retest Date : 04/17/2031

Tested Property	Specification	Analysis
Assay	Min 99.8%	99.94
Color (APHA)	Max 10	5
Residue after evaporation	Max 0.001%	0.0001
Acetone and aldehyde	Passes test	Passes
Acid and chloride	Passes test	Passes
Free chlorine (Cl)	Passes test	Passes
Lead (Pb)	Max 0.05 ppm	<0.05
Substances darkened by sulfuric acid	Passes test	Passes
Stabilizer, ethanol	0.5 - 1.0%	0.8%
Suitability for use in Dithizone Tests	To Pass Tests	Passes
Water	Max 0.03%	0.003

Store in cool, dry, well-ventilated environment.

[Chloroform \$\geq\$ 99% Lab Grade](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2026 Lab Alley



particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2026 Lab Alley