

### Certificate of Analysis

**Product: Deionized Water Lab Grade, Kosher**

**Item Number: WAD**

**CAS#: 7732-18-5**

**Grade: Lab**

**Lot Number: 111025-F**

**Manufacture Date: 11/10/2025**

**Expiration Date: 11/09/2027**

**Country of Origin: Select Country**

Tested Property	Specification	Analysis
Appearance	Clear thin liquid	Pass
Color (APHA)	Max 15 Pt-Co	<15
Odor	Characteristic	Pass
Suspended matter	Substantially free from particulate	Pass
Specific gravity @ 60°F	0.785 - 0.793	0.788
Water Content, % (ppm)	<0.2 (2000)	0.03 (378)
Identity by GC	Compared to standard	Pass
Ethanol, Wt %	Typical Composition	95.8%
n-Heptane, Wt%	Typical Composition	4.2%

#### [Deionized Water Lab Grade, Kosher](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,  
E-mail: [customercare@laballey.com](mailto:customercare@laballey.com),  
[www.laballey.com](http://www.laballey.com),

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2026 Lab Alley