

Certificate of Analysis

Product: Guava ViaCount Reagent (240 mL)

Item Number: 4000-0041

Grade: Reagent

Lot Number: GA4000004161224

Manufacture Date: 11/11/2023

Expiration Date: 11/10/2026

Country of Origin: United States

Tested Property	Specification	Analysis
Total cell count	Benchmark $\pm 10\%$	Pass
% Viability	Benchmark $\pm 10\%$	Pass
Appearance	Faintly reddish-purple solution	Pass
pH	7.15 - 7.25	Pass

Product Description: ViaCount Reagent contains two (2) fluorescent nucleic acid stains in a buffered, physiologically compatible saline solution. Sodium azide is included as a preservative at a concentration of 0.05%.

Test Description: Performance testing on known mixtures of viable and non-viable Jurkat cells, against benchmark lot.

[Guava ViaCount Reagent \(240 mL\)](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,

E-mail: customercare@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2025 Lab Alley