

### Certificate of Analysis

**Product: Isopropyl Alcohol 99% HPLC Grade**

**Item Number: IPAHPLC**

**CAS#: 67-63-0**

**Grade: HPLC**

**Lot Number: 5001906/2.1**

**Manufacture Date: 10/28/2024**

**Expiration Date: 10/27/2027**

**Country of Origin: China**

Tested Property	Specification	Analysis
Purity (GC)	Min 99.9%	99.97
Density (20°C)	0.784 - 0.786 g/mL	0.785
Mixing Test With Water	To Pass Test	Pass
Residue on Evaporation	Max 5 ppm	2.2
Water (by Karl Fischer)	Max 1000 ppm	336.3
Acidity (H+)	Max 0.0003 mmol/g	0.00006
Absorbancy:		
λ: 205nm	Max 0.009	0.003
λ: 220nm	Max 0.032	0.010
λ: 230nm	Max 0.051	0.019
λ: 245nm	Max 0.149	0.055
λ: 260nm	Max 0.301	0.107
λ: 300nm	Max 1.000	0.669

#### [Isopropyl Alcohol 99% HPLC Grade](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This is an electronically generated document and is valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,  
E-mail: [customercare@laballey.com](mailto:customercare@laballey.com),  
[www.laballey.com](http://www.laballey.com),

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2026 Lab Alley