

Certificate of Analysis**Product: Oxalic Acid Crystals****Item Number: OXAGL****Grade: Lab****Lot Number: 5000274/1.2****Manufacture Date: 05/10/2023****Expiration Date: 05/09/2026****Country of Origin: China**

Tested Property	Specification	Analysis
Assay	Min 99.6%	99.72
Sulfate	Max 0.08%	0.07
Chloride	Max 0.001%	0.0004
Ignition 850 °C	Max 0.05%	0.02
Heavy Metals (on Pb basis)	Max 0.001%	0.0006
Iron	Max 0.001	0.0005
Calcium	Max 0.001	0.0005
Sodium	Max 0.001	0.0007
Water Insoluble @25°C	Max 0.1%	0.032
Particle Size	Min 90% passes through 20 mesh	Pass

[Oxalic Acid Crystals](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,

E-mail: customercare@laballey.com,
www.laballey.com,12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2025 Lab Alley