

Certificate of Analysis

Product: Trichloroacetic Acid, 12.5%
Item Number: TCAACLA125
Grade: Lab
Lot Number: 5001266/1.1-125-61925
Manufacture Date: 06/19/2025
Expiration Date: 06/18/2027
Country of Origin: United States

Tested Property	Specification	Analysis
Assay	12 - 13% w/v	12.5%

This material is produced using ACS Trichloroacetic acid with ASTM1193 Deionized Water Type IV.

[Trichloroacetic Acid, 12.5%](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com,

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2025 Lab Alley