LAB ALLEY
ESSENTIAL CHEMICALS

Certificate of Analysis

Product: Water, HPLC Grade
Item Number: WAHPLC
CAS#: 7732-18-5
Grade: HPLC
Lot Number: 5002613/1.1
Manufacture Date: 10/26/2026
Retest Date : 10/25/2029
Country of Origin: United States

Tested Property	Specification	Analysis
Residue after Evaporation	Max 1 ppm	Pass
Liquid Chromatography	To Pass the Test	Pass
Absorbance	To Pass the Test	Pass
UV Absorbance:		
@200nm	Max 0.010 AU	-0.0210 AU
@210nm	Max 0.010 AU	-0.0420 AU
@254nm	Max 0.005 AU	-0.0290 AU
@300nm	Max 0.005 AU	-0.0290 AU
Appearance	Clear	Pass
Heavy Metals (as P)	Max. 0.01 ppm	Pass
Sub. Reducing Permanganate	To Pass the Test	Pass
Chloride	Max. 0.4 ppm	Pass
Phosphate	Max. 1 ppm	Pass
Sulfate	Max. 1 ppm	Pass
Residue after Evaporation	Max. 1 ppm	Pass
UV Spectrophotometry	To Pass the Test	Pass
Gradient Elution	To Pass the Test	Pass

Store in a cool, dry, and well-ventilated environment.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com,
12501 Pauls Valley Road, Suite A, Austin, Texas 78737,
© copyright: 2026 Lab Alley

Water, HPLC Grade

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This is an electronically generated document and is valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com,
12501 Pauls Valley Road, Suite A, Austin, Texas 78737,
© copyright: 2026 Lab Alley