

Certificate of Analysis

Product: Water, LC/MS Grade
Item Number: WALCMS
CAS#: 7732-18-5
Grade: LCMS
Lot Number: 5002659/2.1
Manufacture Date: 06/02/2026
Retest Date : 06/01/2029
Country of Origin: United States

Tested Property	Specification	Analysis
Residue after Evaporation	Max 1 mg/L	0.2 mg/L
UV Absorbance Test	To Pass the Test	Pass
Refractive Index @ 20°C	1.3325 - 1.3335	1.3330
LC Gradient elution suitability	To Pass the Test	Pass
Fluorescence at emission max. for imp.	Max 1.0 ppb as Quinine sulfate	<0.3 ppb
LC-MS Positive mode (as Reserpine)	Max 50 ppb	<50 ppb
LC-MS Negative mode (as Nitrophenol)	Max 50 ppb	<50 ppb
Trace Metals	To Pass the Test	Pass

Store in dry, cool well-ventilated environment.

[Water, LC/MS Grade](#)

Note: The information and recommendations of Lab Alley concerning this product are based upon laboratory tests and experience. To the best of our knowledge and belief these are true and accurate, however Lab Alley assumes no obligation or liability for the information in this document. Since conditions of actual use are beyond our control, any recommendations or suggestions regarding merchantability and fitness for particular purposes are made without warranty, expressed or implied.

This is an electronically generated document and is valid without a signature.

Phone: 512-668-9918, Fax: 512-886-4008,
E-mail: customercare@laballey.com,
www.laballey.com.

12501 Pauls Valley Road, Suite A, Austin, Texas 78737.

© copyright: 2026 Lab Alley