



TECHNICAL DATA SHEET(TDS)

Water, LC/MS Grade

Product Specifications (CAS# 7732-18-5)

Test	Specification	Typical Results
Residue after Evaporation	Max 1mg/L	<1mg/L
LC Gradient elution suitability	To Pass the Test	Pass
Fluorescence at emission max. for imp.	Max 1.0 ppb as Quinine sulfate	< 0.3 ppb
LC-MS Positive mode (as Reserpine)	Max 50 ppb	< 50 ppb
LC-MS Negative mode (as Nitrophenol)	Max 50 ppb	< 50 ppb
Trace Metals	To Pass the Test	Pass
UV Absorbance Test	To Pass the Test	Pass

Appropriate/legal use of this product is the responsibility of the user.

[Water, LC/MS Grade](#)

Phone: 512-668-9918, Fax: 512-886-4008, E-mail: customercare@laballey.com
www.laballey.com

12501 Pauls Valley Road, Suite A, Austin, Texas 78737
© 2026 Lab Alley