



TECHNICAL DATA SHEET(TDS)

Water, USP Grade

Product Specifications (CAS# 7732-18-5)

Test	Specification	Typical Results
Appearance @ 25°C	Clear, colorless liquid	Pass
Water Conductivity (Point of fill)	NMT 1.3 μ S/cm at 25°C	0.05
Microbial Limits	NMT 100 CFU/mL	1
Total Organic Carbon	NMT 0.50 mg/L	0.24

Appropriate/legal use of this product is the responsibility of the user.

[Water, USP Grade](#)

Phone: 512-668-9918, Fax: 512-886-4008, E-mail: customercare@laballey.com
www.laballey.com

12501 Pauls Valley Road, Suite A, Austin, Texas 78737
© 2026 Lab Alley