



TECHNICAL DATA SHEET(TDS)
Xylenes Histological Lab Grade
Product Specifications (CAS# 106-42-3)

Test	Specification	Typical Results
Appearance	Clear and Bright	Pass
Assay	>98.50%	99.08%
Sulfur	Max 1.0 mg/kg	<0.5 mg/kg
Water	Max 0.05%	0.0112%
Acidity	As reported	No free acid
Color Pt-Co	Max 20	3
Specific Gravity @ 60°F	As reported	0.8713

Appropriate/legal use of this product is the responsibility of the user.

[Xylenes Histological Lab Grade](#)

Phone: 512-668-9918, Fax: 512-886-4008, E-mail: customercare@laballey.com
www.laballey.com

12501 Pauls Valley Road, Suite A, Austin, Texas 78737
© 2026 Lab Alley